

Glycated albumin assay kit(GA-D)



Each Molecular Promises Accuracy
info@hzymes.com
www.hzymes.com

Manual Version: V1.0

Glycated albumin assay kit(GA-D)

Catalog number

HOEM1008-01

HOEM1008-02

HOEM1008-03

HOEM1008-04

Applications^[1,2]

Used for quantitative determination of glycated albumin (GA) and albumin (ALB) in human serum, the ratio of glycated albumin (%) is used to determine glycated albumin concentration divided by the albumin concentration.

Glycated albumin is a non-enzymatic reaction between glucose and albumin in the blood, because the half-time of albumin in vivo is short (about 17-19 days), so GA serves as a 2-3 week indicator of average blood glucose. Therefore, it has significant clinical application value. When blood sugar needs to be evaluated for short-term control.

Principle

1. Determination of glycated albumin

In the sample (serum), specific proteases are added to decompose Glycated albumin to form glycosylated amino acids. Then glycosylated amino acids are converted into glucoaldosterone, amino acids, and hydrogen peroxide by specific glycosylated amino acid oxidase. Under the action of peroxidase, hydrogen peroxide reacts with 4-amino-antipyrine and TOOS to form red quinone-imine^[3]. The concentration of glycated albumin in the sample was obtained by measuring its absorbance.

2. Determination of albumin

In pH 4.2, the albumin in the sample (serum) binds with the indicator bromocresol green to form a blue-green complex. The concentration of albumin in the sample was obtained by measuring the change in absorbance.

3. Calculation of glycated albumin value (%)

$$\text{Glycated albumin(\%)} = \frac{\text{GA (g/L)}}{\text{ALB (g/L)}} \times 100\%$$

The albumin reagent originated from Hzymes.

Reagents

Reagent	Components	Concentrations
GA	Reagents 1 (R1) :	
	ADA buffer	20mmol/L
	PRK	400KU/L
	HTBA	10mmol/L
	Reagents 2 (R2) :	
	FAOD	100KU/L
	Peroxidase	20KU/L
ALB	4-Aminoantipyrine	10 mmol/L
	Reagents 1 (R1)	
	Succinic acid buffer	120mmol/L
	Tween 80	0.1%
	Reagents 2 (R2)	
	Succinic acid buffer	120mmol/L
	Bromocresol purple	0.15mmol/L

Sample requirements

1. Serum, heparin or EDTA plasma, and urine are suitable for samples. Whole blood, hemolysis is not recommended for use as a sample. Freshly drawn serum is the preferred specimen.
2. Stability: Serum/plasma: 7 days at 2-8°C.

Calibrator preparation

Carefully open the bottle, avoiding the loss of lyophilizate, and pipette in exactly

1.0 mL of distilled/deionized water. Carefully close the bottle and dissolve the contents completely by occasional gentle swirling within 30 minutes. Avoid the formation of foam. The dissolved calibrator can be used without any other pretreatment.

Quality control preparation

Carefully open the bottle, avoiding the loss of lyophilizate, and pipette in exactly 1.0 mL of distilled/deionized water. Carefully close the bottle and dissolve the contents completely by occasional gentle swirling within 30 minutes. Avoid the formation of foam. The dissolved control can be used without any other pretreatment.

Method

1. Reagent preparation : Liquid reagent can be used when opened
2. Measurement:
 - 1) GA:

Main wavelength	600nm	Sub wavelength	700nm
Temperature	37°C	Type	2-End Point
Sample (calibration)		16μL	
R1		160μL	
Mix, incubated at 37°C for 5min to determine the absorbance of A1.			
R2		40μL	
Mix, incubated at 37°C for 5min to determine the absorbance of A2.			

$$\Delta A = [(A2-A1)_{\text{sample}}] - [(A2-A1)_{\text{blank}}]$$

2) ALB:

Main wavelength	600 nm	Sub wavelength	660 nm
Temperature	37 °C	Type	2-End Point
Sample (calibration)	3µL		
R1	220µL		
Mix, incubated at 37 °C for 4.5min to determine the absorbance of A1.			
R2	110µL		
Mix, incubated at 37 °C for 9min to determine the absorbance of A2.			

$$\Delta A = [(A2-A1)_{\text{sample}}] - [(A2-A1)_{\text{blank}}]$$

Calibration

It is recommended to use the Calibrator from Hzymes and distilled/deionized water for two-point calibration. If the reagent changes the batch, recalibration should be performed.

Quality control

At least two levels of control material should be analyzed with each batch of samples. Each laboratory should establish its own internal quality control scheme and procedures for corrective action if controls do not recover within the acceptable tolerances.

Reference intervals^[3]

Each laboratory should establish its own reference intervals based on its patient population. The reference intervals measured at 37°C listed below were taken from the literature.

Serum / Plasm: 10.8%~17.1%

Interferences/Specificity

The following substances were tested for interference with this methodology. Criterion: Recovery within ± 10% of the initial value.

Substance	Level Tested	Observed Effect
Lipemia	100 mg/dL	NSI
Hemoglobin	500 mg/dL	NSI
* NSI: No Significant Interference (within ± 10%)		

Storage and stability

Up to expiration date indicated on the label, when stored unopened at 2-8°C and protected from light.

Once opened, the reagents are stable for 30 days when refrigerated on the analyzer or refrigerator.

Contamination of the reagents must be avoided. Do not freeze the reagents.

Once dissolved, the calibrator are stable for 7 days at 2-8°C, the control are stable

for 7 days at 2-8°C, do not freeze.

Reagent blank absorbance

The Blank absorbance of the GA reagent should be ≤0.1A. The Blank absorbance of the ALB reagent should be ≤0.35.

Analytical sensitivity

When the sample concentration for testing is 1.0 g/dL, the absorbance difference before and after the reaction should not be less than 0.0536.

When the concentration of ALB is 4.11 g/dL, the absorbance difference before and after the reaction should not be less than 0.2;

Precision

Within-run: CV ≤ 5%

Between-run: CV ≤ 10%

Linearity range

Conventional Units:

GA/ALB: 6.4%~63.3%

The concentration of the test sample should be within the linear range. If the value of sample exceeds 63.3%(GA/ALB), the sample should be diluted with 9 g/L NaCl solution.

Warnings and precautions

1.For in vitro diagnostic use.

2.Take the necessary precautions for the use of laboratory reagents.

3.Preservative contained. Do not swallow. Avoid contact with skin and mucous membranes.

4.Disposal of all waste material should be in accordance with local guidelines.

5.Material safety data sheet is available on request for professional users.

References

[1] Guide for clinical application of blood glucose monitoring in China (2011 edition). Chinese Journal of Diabetes , 2011, 3(1): 13 -21.

[2] Santiago Rodriguez-Segade et al. Progression of nephropathy in type 2 diabetes: The glycation gap is a significant predictor after adjustment for glycohemoglobin (HbA1c). Clinical Chemistry, 2011, 57(2):264-271.

[3] M. Koga, et al. Glycated albumin and glycated hemoglobin are influenced differently by endogenous insulin secretion in patients with type 2 diabetes. Diabetes care, 2010, 33(2): 270 – 272.