



## Certificate of Analysis

<b>Product Name</b>	Glycated albumin (GA) Quality control 1		
<b>Catalog Number</b>	HH-QC 1202-01		
<b>Batch Number</b>	HH20220104	<b>Expiring date</b>	18 months
<b>Recommended Storage</b>	2-8°C		
<b>Tests Item</b>	<b>Specifications</b>	<b>Results</b>	
<b>Appearance</b>	White freeze-dried powder	PASS	
<b>Difference between bottles</b>	≤5%	PASS	
<b>Water content</b>	≤6%	PASS	
<b>Re solution stability at 2-8 °C</b>	After the product is re dissolved with water, it is stored at 2-8 °C and the absolute deviation of the test value deviation is ≤± 10% within 7 days	PASS	
<b>Stability at 37 °C</b>	The product is stored under closed cover at 37 °C, and within 7 days, the absolute deviation of the detected value deviation is ≤ ± 10%	PASS	
<b>Conclusion</b>	Pass		
<b>Analyst</b>	张丹 (QC)	<b>Date</b>	2022.01.13
<b>Corrector</b>	孙湘云 (QA)	<b>Date</b>	2022.01.13

