



Certificate of Analysis

Product Name	Glycated albumin (GA) Quality control 1		
Catalog Number	HH-QC 1202-02		
Batch Number	HH20220104	Expiring date	18 months
Recommended Storage	2-8°C		
Tests Item	Specifications	Results	
Appearance	White freeze-dried powder	PASS	
Difference between bottles	≤5%	PASS	
Water content	≤6%	PASS	
Re solution stability at 2-8 °C	After the product is re dissolved with water, it is stored at 2-8 °C and the absolute deviation of the test value deviation is ≤± 10% within 7 days	PASS	
Stability at 37 °C	The product is stored under closed cover at 37 °C, and within 7 days, the absolute deviation of the detected value deviation is ≤ ± 10%	PASS	
Conclusion	Pass		
Analyst	张丹 (QC)	Date	2022.01.13
Corrector	孙湘云 (QA)	Date	2022.01.13

